

Ing. Christian Fuczik Chemisches Labor GmbH Gerhardusgasse 25/3. OG, 1200 Wien E-Mail: info@hanfanalytik.at Tel.: +43 660 867 0063 www.hanfanalytik.at

## Certificate of Analysis Cannabinoids

Ketama CBG Description I: Client: 31/07/2023 Sample ID: E9300002 Sample date: Sample material: Bloomday: resin

Description II: Further information: Lotto 19

Abbr.	Cannabinoids Basic	Result	Unit
T-CBD	Total Cannabidiol (CBD + CBDA)	10.53	% (w/w)
CBD	Cannabidiol	10.01	% (w/w)
CBDA	Cannabidiolic acid	0.59	% (w/w)
T-THC	Total Tetrahydrocannabinol (THC + THCA)	0.07	% (w/w)
D9THC	D9-Tetrahydrocannabinol	0.07	% (w/w)
THCA	Tetrahydrocannabinolic acid	ND**	% (w/w)
D8THC	D8-Tetrahydrocannabinol	ND**	% (w/w)
T-CBG	Total Cannabigerol (CBG + CBGA)	15.89	% (w/w)
CBG	Cannabigerol	1.10	% (w/w)
CBGA	Cannabigerolic acid	16.85	% (w/w)
CBN	Cannabinol	ND**	% (w/w)
CBC	Cannabichromene	0.22	% (w/w)
CBDV	Cannabidivarin	ND**	% (w/w)
CBDVA	Cannabidivarinic Acid	ND**	% (w/w)
THCV	Tetrahydrocannabivarin	ND**	% (w/w)

Sample received: 01/08/2023 - 3,801 g



Head of Laboratory Services

Ing. Christian Fuczik, Chemist Analysis reviewed - last changes: 03/08/2023 at 11:01

\*\*) ND =not detectable. The measured value was below the limit of detection of 0.01% or 100 mg/kg. The expected measurement uncertainty varies with substance and concentration and can be assumed to be a maximum of 10 %.

For the calculations of the equivalent sums, the respective acid forms were multiplied by the factor 0.877 or 0.878 to conclude the equivalent amount of the neutral

Method of analysis: HPLC-DAD (High Performance Liquid Chromatography - Diode Array Detector) according to Ph.Eur. 2.2.29 (European Pharmacopoeia) This Certificate of Analysis may only be reproduced as a whole and not in parts. Any alteration is punishable under § 223 StGB (Austrian Penal Code) (forgery of







