



CERTIFICATE OF ANALYSIS

Sample Number

2026A4208

Sample Data

Name: BUBBLE
Lot: 42080226
Customer:
Sampling: Performed by the customer
Batch Result: Pass
Moisture: N/C

Analysis Data

Delivery Date: 12.07.2026
Starting Date: 12.07.2026
Ending Date: 12.07.2026
Analysis Method: HPLC-DAD Analysis
Sample Preparation: Ultrasound-assisted extraction with methanol-chloroform

Results

Results expressed in percentages.

<u>ANALYTE</u>	<u>LABELED AMOUNT</u>	<u>DETECTED</u>	<u>LOQ</u>	<u>LOD</u>
Δ9-THC		< N/D%	0,01%	0,005%
THCA		0,124% ± 0,00%	0,01%	0,005%
CBD		1,244% ± 0,01%	0,01%	0,005%
CBDA		4,975% ± 0,04%	0,01%	0,005%
CBDV		< N/D%	0,01%	0,005%
CBDVA		< N/D%	0,01%	0,005%
THCV		< N/D%	0,01%	0,005%
THCVA		< N/D%	0,01%	0,005%
CBG		< N/D%	0,01%	0,005%
CBGA		< N/D%	0,01%	0,005%
CBC		< N/D%	0,01%	0,005%
CBCA		< N/D%	0,01%	0,005%
CBN		< N/D%	0,01%	0,005%
CBNA		< N/D%	0,01%	0,005%
Δ8-THC		< N/D%	0,01%	0,005%
CBL		< N/D%	0,01%	0,005%

Sample Number
2026A4208

Total results obtained using a theoretical conversion factor.

THC total	0,109% ± 0,001%	0,01% 0,005%
CBD Total	5,607% ± 0,044%	0,01% 0,005%
CBDV Total	< N/D%	0,01% 0,005%
THCV Total	< N/D%	0,01% 0,005%
CBG Total	< N/D%	0,01% 0,005%
CBC Total	< N/D%	0,01% 0,005%
CBN Total	< N/D%	0,01% 0,005%
COMMENTS		

Die Ergebnisse dieser Analyse entsprechen der Analyse der vom Kunden gelieferten Probenmenge und können von der Gesamtchargenmenge abweichen. Der Kunde ist dafür verantwortlich, dass die gelieferte Probe nicht repräsentativ für die gesamte Charge ist. Der Bericht dient nur zu Informationszwecken. Sie darf unter keinen Umständen zur Diagnose, Behandlung oder Prävention von Krankheiten oder Gesundheitsproblemen verwendet werden. Kein Teil dieses Dokuments darf ohne Genehmigung des Schweizerischen Kristalllabors reproduziert werden.

Verantwortlicher Analytiker:



M. Martinez

Analytics Laborleiter:



A. Cantalapiedra

Swiss Cristal Lab AG / Labor, Badische Bahnhofstrasse 16 / CH-8212 Neuhausen am Rheinfall.

m.martinez@swiss-cristal-lab.ch